

SECRET

NPIC/TSSG/DED-1127-68
18 March 1968

Declass Review by NGA.

MEMORANDUM FOR THE RECORD

SUBJECT: Contract Negotiation with [REDACTED]
[REDACTED] for Fabrication of a Rapid Alignment Device for Microstereo-
scope

1. Changes to the Technical Requirements and Performance Specifications for the Rapid Alignment Device follow. These changes were the result of discussions during a conference telephone call on 8 March 1968. Conferees were

[REDACTED]
[REDACTED] of the Agency.

2. Paragraph 3.1.1. Overall dimensions of the device were changed to 7 inches wide, 4 1/2 inches high, and 4 inches deep. The drawing on page 2-1c of [REDACTED] Technical Proposal WDP-232 for the device is to scale and indicates the required dimensions of [REDACTED] design.

3. Paragraph 3.2.1.1. Resolution shall not be reduced by more than 20%.

4. Paragraph 3.2.1.2. This requirement was questioned as to necessity for limiting magnitude to 60%, and as to determination and measurement of the amount of reduction. The exact requirement will be established after further discussion.

5. Paragraph 3.3.2.1. First sentence will read, "Where reflective optics are incorporated within the system, they shall not be subject to realignment due to reasonable handling."

6. Paragraph 3.4.1. The instruction and maintenance books will be type-written sheets and need not be formal bound manuals.

7. Paragraph 3.4.3. Drawings will consist of an assembly to scale and details in the form of sketches. Correct dimensions will appear on these sketches but they may not be formal drawings.

TSSG/DED/RQDB

Distribution:

Orig - Rt & File
2 - Chronos

SECRET

